



COUPE A-A

Checkvalve in stand by position:	Closed
Maximum retrograde leakrate:	≤ 0.1 ml/h at a flow rate of 0.5ml/h
Position of use:	Independent
Priming:	Airbubbles completely removable
Crack pressure:	≤ 6 mbar
Crack pressure after back pressure of 6 bar for 15 min:	≤ 12 mbar
Maximum back pressure:	8 bar
Burst pressure (at room temperature in flow direction for 15 seconds air overpressure):	8 bar
Flowrate with infusion set according to DIN 8536 with Glucose 40%:	≥ 110 ml/ min
Flowrate with infusion set according to DIN 8536 with water (0.9% NaCl):	≥ 150 ml/ min
Sterilization :	ETO-Gas, γ -irradiation up to 50 kGy

Rep	Matière	Code matière
1	ABS +Colorant blanc	M0480+M0507
2	ABS transparent	M0480
3	Silicone/ Elastosil	M0256

ORIGINAL

APPROBATION / APPROVAL		
Client / Customer :	Promepla Resp. BE : 17.12.2013 DANIEL J	Promepla Qualité : 21 FEB. 2014 D. AB
Nom / Name, date & signature		

1	03/05/2013	RANDAZZO	Mise à jour désignation, cotation, spécification technique. Code matière M0071 devient M0480 et M0114 devient M0256. Mise au format SW.	FOR021: 4874
0	22/02/2008	RANDAZZO	Creation	
Rev.	Date	Designer	Modifications	N° Fiche FOR073 ou FOR021

Norms :			
Material : VOIR NOTA	P. 1 / 1	Scale : 5:2	
Material code : VOIR NOTA	Revision : 1	Format : A4	
Reference : PBP21053	General tolerancias :	Units : mm, °, grs	
Designation : VALVE ABS - PRESSION OUVERT. 6Mbar - TUBE Ø 3x4.1mm - Ø 3x4.1mm		PROMEPLA	
		FOR035-1	